



WASHINGTON STATE PATROL  
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: 4/24/2006 INITIALS: DEC OWNER: WSP  
DISTRICT: 1 ADDRESS: \_\_\_\_\_  
LOCATION: 7th Shop COUNTY: \_\_\_\_\_  
MAKE: Hanni TYPE: port wheel weigher ACCURACY CLASS: III  
MODEL: 101 ID. NO: 24083 CAPACITY: 20,000 SECTIONS: 1  
MINIMUM GRADUATIONS: 50 REG. ELEMENT: dial PLATFORM: 15" X 26"


LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
Center	5000	5000	0				
"	10000	10050	+50				
"	15000	15050	+50				
Center	20000	20000	0				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as is shown on the attached "Certificate of Accuracy", No. 19577

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cronin Date: 4/26/2006

State of Washington  
County of Thurston  
Signed and sworn to (or affirmed) before  
me on 4-26-06 by  
Michael J. Evans  
(Signature of notary public)  
MICHAEL J. EVANS  
Printed Name of Notary  
My appointment expires 12-27-07



3000-362-021 (R) (10/98)

DISTRIBUTION: WHITE - COURT GREEN - DISTRICT CANARY - PROPERTY MANAGEMENT PINK & GOLDENROD - SERVICE TRUCKS